

CERTIFICATE OF COMPLIANCE

Certificate Number 20130514-E191327
Report Reference E191327-20130513
Issue Date 2013-May-14

Issued to: MAIN POWER ELECTRIC CO LTD,
64 TAO TE RD, PA TE,
TAOYUAN HSIEN 334 TAIWAN.

**This is to certify that
representative samples of**

SYSTEMS, ELECTRICAL INSULATION
Systems, Electrical Insulation- Class 155 (F) transformer
insulation system designated MP SAF004.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.



Standard(s) for Safety:

UL 1446 STANDARD FOR SYSTEMS OF INSULATING
MATERIALS – GENERAL.
CSA C22.2 NO. 0-10 GENERAL REQUIREMENTS --
CANADIAN ELECTRICAL CODE.

Additional Information:

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

